



बैंक नोट पेपर मिल इण्डिया प्रा. लिमिटेड

BANK NOTE PAPER MILL INDIA PVT LIMITED

JV of SPMCIL - A Govt. of India Enterprises & BRBNMPL - A Subsidiary of RBI

**BANK NOTE PAPER MILL INDIA PRIVATE LIMITED**

**CIN: U21090KA2010PTC055475**

**CORPORATE OFFICE, MYSORE**

[gopalakrishnan.m@bnpmindia.com](mailto:gopalakrishnan.m@bnpmindia.com)

**Ph: 0821 2401 158**

**Limited Tender Enquiry No: BNPM/LTE/AIRFLOWROTAMETER/325/2018-19**

**Issuing Date: 17.07.2018 11:00 hrs.**

**Closing date: 31.07.2018 15:00 hrs.**

**1. Scope of work:**

Supply of Air Flow Rotameter as per given technical specifications.

**2. Terms & Conditions:**

S. No	Particulars	Quantity
1.	Payment Terms	100% on receipt and acceptance of goods by the consignee at destination and on production of all required documents by the supplier.
2.	Price	Price should be inclusive of all taxes & duties.
3.	Freight	Shall be included in quoted price.
4.	Packing & Forwarding	Shall be included in quoted price.
5.	Delivery Terms	F.O.R, Bank Note Paper Mill India Private Limited, Mysore
6.	Warranty Period	.....
7.	Warranty Certificate	Applicable / Not Applicable
8.	Delivery Period	.....
9.	Material to be delivered at	Engineering Stores, Bank Note Paper Mill India Private Limited, Note Mudran Nagar, Mysore 570 003.
10.	Contact person	Mr. Gopalakrishnan M P: 0821-2401158




**Technical Specifications:**

Air Flow Rotameter

- a. Flow Range: 25 – 250 LPM.
- b. Connection: ¼ inch NPT (F).
- c. Orientation: Rear to Rear.
- d. Body: Acrylic Plastic Body as clear as glass.
- e. Scale: Calibrated Scale.
- f. Mounting: Wall mounting (along with KIT).
- g. MOC of wetted parts: 316 SS (or) better.
- h. Packing: Teflon / PTFE.
- i. Temp: Up to 45 Deg C Max.
- j. Accuracy: +/- 2% of FR (or) better.
- k. Air pressure: 6 bar.
- l. Knob: Required.
- m. Float System: Guided Rod Type Float.

**Drawing:**



**Limited Tender Enquiry No: BNPM/LTE/AIRFLOWROTOMETER/325/2018-19****Quality Control Parameters:**

<b>S No</b>	<b>Technical Parameters</b>	<b>BNPM requirements</b>	<b>Bidder offered specification</b>	<b>Deviations/Remarks (if any)</b>
1	Material	Air Flow Rotameter		
2	Flow Range	25 – 250 LPM		
3	Scale units	LPM		
4	Connection size	¼ inch NPT (F)		
5	Connection orientation	Rear to Rear		
6	Flow meter body	Acrylic plastic body as clear as glass		
7	Scale	Calibrated scale engraved on acrylic body itself		
8	Metering Tube	Drilled and reamed in Acrylic itself		
9	Mounting	Wall mounting kits required		
10	MOC of wetted parts (Float & Fittings)	SS 316 (or) better		
11	Packing	Teflon/PTFE		
12	Temp rating	Up to 45°C Max		
13	Accuracy	+/- 2% F.R (or) better		
14	Air pressure available	6 Bar		
15	Knob operated flow control	Required		
16	Float System	Guided rod type float mechanism is preferable		

(To be filled, signed and stamped by bidder and to be sent along with price bid)



**Limited Tender Enquiry No: BNPM/LTE/AIRFLOWROTOMETER/325/2018-19**

**Price Bid Format :**

S No	Description	Qty	UOM	Unit rate	Freight, P&F Charges	GST @ 18% (HSN - 9026)	Unit rate including GST & Freight, P&F Charges	Grand total inclusive of GST and Freight, P&F Charges (D.A.P, BNPM, Mysore)
1	Supply of Air Flow Rotameter as per given technical specifications	2	Nos					
<b>Total price inclusive of all tax, freight , P&amp;F charges</b>								

**HSN Codes:** (90261010)

We hereby confirm that

1. We accept all terms & conditions mentioned in the enquiry.
2. Price quoted is inclusive of all taxes, P&F, freight etc. on D.A.P, BNPM, Mysore basis.

Signature of bidder:.....

Name of the Firm:.....

Seal of the firm: .....

GST No: .....

