

BANK NOTE PAPER MILL INDIA PRIVATE LIMITED CIN: U21090KA2010PTC055475 | CORPORATE OFFICE, MYSORE

vaishnavi.shankar@bnpmindia.com

Ph: 0821 2401 175

Limited Tender Enquiry No: BNPM/LTE/Digital Sound Level Meter/116/2019-20

Issuing Date & Time: 08.05.2019 10:00 hrs. Closing Date & Time: 15.05.2019 10:00 hrs.

1. Scope of work:

i. Digital Sound Level Meter-1 No.

- Measurement Range: 30-130db

- Accuracy: +/- 1.5db

- Frequency: 31.5 Hz to 8 KHz

- Other Features: Peak Hold

- Display: LED Display

- Measurement Memory: max. 30 readings

- Power Supply: Battery operated

- Equipment should contain one wind suppressor (foam ball) and operating manual

- Warranty: 24 months

- Initial Calibration Certificate required along with equipment

2. Terms & Conditions:

S. No	Particulars	Quantity
1.	Payment Terms	100% within 30 days from date of receipt and acceptance of goods by the consignee at destination and on production of all required documents by the supplier.
2.	Price	Price should be inclusive of all taxes & duties.
3.	Freight	Shall be included in quoted price.
4.	Packing & Forwarding	Shall be included in quoted price.
5.	Delivery Terms	F.O.R – Bank Note Paper Mill India Private Limited, Mysore.
6.	Warranty Period	from date of supply at BNPM, Mysuru.
7.	Warranty Certificate	Applicable / Not Applicable
8.	Delivery Period	weeks from date of receipt of purchase order.



9.	Validity of Bid	30 days from date of closing of tender.	
10.	Material to be delivered at	Engineering Stores, Bank Note Paper Mill India Private Limited, Note Mudran Nagar, Mysore 570 003.	
11.	Contact person	Ms. Vaishnavi Shankar P: 0821-2401175	



Price Bid Format:

S No	Description	Qty	UOM	Unit rate	Freight, P&F Charges	GST @ 18%	Unit rate includin g GST & Freight, P&F Charges	Grand total inclusive of GST and Freight, P&F Charges (F.O.R, BNPM, Mysore)
	Digital Sound Level Meter							
	Measurement Range: 30-							
	130db							
	Accuracy: +/- 1.5db							
	Frequency: 31.5 Hz to 8							
	KHz							
	Other Features: Peak Hold							
	Display: LED Display Measurement Memory:							
	Measurement Memory: max. 30 readings							
1.	Power Supply: Battery	1	No.					
1.	operated							
	Equipment should contain							
	one wind suppressor							
	(foam ball) and operating							
	manual							
	Warranty: 24 months							
	Initial Calibration							
	Certificate required along							
	with equipment							
	Total price inc	.la!	a f a ll + -	<i>c</i>	DOF O	41	•	

TION		•	-
$H \setminus V$	Code:	1	
11.714	GOUC.		



We hereby confirm that

- 1. We accept all terms & conditions mentioned in the enquiry.
- 2. Price quoted is inclusive of all taxes, P&F, freight etc. on F.O.R, BNPM, Mysore basis.

Signature of bidder:
Name of the Firm:
Seal of the firm:
GST No:

